

LAB ANALYSIS
W91ZRS-04-Q-0009
AMENDMENT 2

June 14, 2004

This solicitation is amended to provide the following information:

1. Is the lab required to hold a current Army Corps of Engineers validation?

Answer to Question 1 - **NO**

2. Regarding method 525.2 for semi volatile analysis, can the lab substitute method 8270 SIM as long as detection limits can be achieved?

Answer to Question 2 - **YES**

3. Method 8080 is listed for Pest/PCB. This method has been deleted from SW-846 and replaced by methods 8081(pest) & 8082 (pcb). Can I assume that this will be allowed?

Answer to Question 3 - **YES**

4. For Arsenic & Selenium on liquids furnace methods are indicated. Method 6010 with appropriate correction is an acceptable method for As & Se. Can the lab substitute 6010? (method 6010 is indicated as acceptable for Se on soils in your tables)

Answer to Question 4 - **YES**